

PolyTechnic Resources, Inc.

P.O. Box 1557
198 Venture Drive
Seaford, DE 19973
Phone: 302-629-4221 * Fax: 302-629-3541

Statement of Cross-Contamination Efficiency of PFT Filters

This is to certify that PFT Filters manufactured by Polytechnic Resources used for Spirometry and Pulmonary Function Testing have been tested by Nelson Labs at an Increased Challenge Level and have shown a Patient Cross-Contamination Prevention Efficiency of greater than 99.999% against Viral and greater than 99.999% against Bacterial Bioburden.

The risk of Patient Cross-Contamination is greatly reduced beyond these efficiencies when the test equipment is properly sanitized between patients, according to the test equipment manufacturer's guidelines.

Timothy C. Byrd
President

The above Calculated Cross-Contamination Prevention data is based on actual test results of PFT Filters. No equipment was needed for the calculations. Ref. Nelson Labs Test Report # 668887 and #668888

Nelson Laboratories
6280 South Redwood Road
Salt Lake City, Utah 84123-6600